

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030284.D
 Acq On : 28 Jul 2022 20:53
 Operator : JC/MD
 Sample : N3898-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 61422-B

Quant Time: Jul 29 06:00:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	252658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	148756	51.475	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.960%
35) Dibromofluoromethane	5.397	113	120168	48.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.560%
50) Toluene-d8	8.653	98	452087	47.391	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.780%
62) 4-Bromofluorobenzene	11.085	95	162288	45.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	6688	9.488	ug/l	98
16) Acetone	2.380	43	5627	3.803	ug/l #	88
19) Methyl tert-butyl Ether	3.123	73	2274	0.296	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
Data File : VX030284.D
Acq On : 28 Jul 2022 20:53
Operator : JC/MD
Sample : N3898-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
61422-B

Quant Time: Jul 29 06:00:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

