

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030285.D
 Acq On : 28 Jul 2022 21:16
 Operator : JC/MD
 Sample : N3695-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-03-07112022

Quant Time: Jul 29 06:00:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	246197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	142617	51.923	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.840%
35) Dibromofluoromethane	5.391	113	116988	49.708	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.420%
50) Toluene-d8	8.653	98	427958	46.954	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.900%
62) 4-Bromofluorobenzene	11.085	95	154363	45.635	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.260%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12032	8.556	ug/l	89
25) 2-Butanone	4.580	43	4065	1.720	ug/l	90
52) Toluene	8.726	92	1445	0.222	ug/l	99
64) Tetrachloroethene	9.275	164	3082	1.162	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
Data File : VX030285.D
Acq On : 28 Jul 2022 21:16
Operator : JC/MD
Sample : N3695-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-03-07112022

Quant Time: Jul 29 06:00:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

