



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072822\
 Data File : VX030264.D
 Acq On : 28 Jul 2022 10:52
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Title : SW846 8260
 Last Update : Mon Jul 25 16:49:17 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	13753	PASS
75	95	30	60	50.6	35736	PASS
95	95	100	100	100.0	70640	PASS
96	95	5	9	6.4	4497	PASS
173	174	0.00	2	1.2	546	PASS
174	95	50	100	64.3	45392	PASS
175	174	5	9	7.2	3284	PASS
176	174	95	101	97.0	44021	PASS
177	176	5	9	6.5	2883	PASS