

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072823\
 Data File : VX036696.D
 Acq On : 28 Jul 2023 10:46
 Operator : JC/MD
 Sample : VX0728WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0728WBL01

Quant Time: Jul 28 16:16:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X071423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 14 04:32:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	31198	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	182656	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	159357	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	76116	29.396	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.000%
60) 4-Bromofluorobenzene	11.079	95	75483	29.201	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.333%
63) Toluene-d8	8.647	98	217144	29.612	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072823\
Data File : VX036696.D
Acq On : 28 Jul 2023 10:46
Operator : JC/MD
Sample : VX0728WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0728WBL01

Quant Time: Jul 28 16:16:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X071423W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 14 04:32:22 2023
Response via : Initial Calibration

