

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
 Data File : VX042613.D
 Acq On : 26 Jul 2024 21:12
 Operator : JC/MD
 Sample : P3334-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E20H1

Quant Time: Jul 27 01:36:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jul 27 01:23:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	159994	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	138628	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	51264	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	37450	38.528	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	77.060%	
7) Chloroethane-d5	1.648	69	17941	44.792	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.580%	
11) 1,1-Dichloroethene-d2	2.294	65	18448	38.520	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	77.040%	
21) 2-Butanone-d5	4.471	46	72513	100.654	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	=	100.650%	
24) Chloroform-d	5.056	84	95372	44.827	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.660%	
26) 1,2-Dichloroethane-d4	5.958	65	64518	47.004	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.000%	
32) Benzene-d6	5.970	84	187998	48.864	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.720%	
36) 1,2-Dichloropropane-d6	7.306	67	57866	50.819	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.640%	
41) Toluene-d8	8.647	98	156498	45.367	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.740%	
43) trans-1,3-Dichloroprop...	8.952	79	23066	43.287	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	86.580%	
47) 2-Hexanone-d5	9.384	63	51724	101.599	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	101.600%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	83299	50.808	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.620%	
66) 1,2-Dichlorobenzene-d4	12.323	152	49968	51.576	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	103.160%	
Target Compounds						
					Qvalue	
13) Acetone	2.392	43	3050	5.416	ug/L	98
20) cis-1,2-Dichloroethene	4.495	96	3240	2.672	ug/L	90
34) Trichloroethene	7.129	95	37896	33.435	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
Data File : VX042613.D
Acq On : 26 Jul 2024 21:12
Operator : JC/MD
Sample : P3334-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E20H1

Quant Time: Jul 27 01:36:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jul 27 01:23:37 2024
Response via : Initial Calibration

