

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
 Data File : VX042648.D
 Acq On : 27 Jul 2024 19:36
 Operator : JC/MD
 Sample : P3373-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E08K0

Quant Time: Jul 29 02:46:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jul 29 02:34:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	177747	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	160936	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	60933	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	34605	32.045	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.100%
7) Chloroethane-d5	1.648	69	17962	40.365	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.740%
11) 1,1-Dichloroethene-d2	2.300	65	20190	37.947	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.900%
21) 2-Butanone-d5	4.465	46	67614	84.480	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.480%
24) Chloroform-d	5.056	84	95739	40.505	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.000%
26) 1,2-Dichloroethane-d4	5.958	65	63396	41.574	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.140%
32) Benzene-d6	5.971	84	187307	41.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.880%
36) 1,2-Dichloropropane-d6	7.312	67	57308	43.352	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.700%
41) Toluene-d8	8.647	98	163334	40.786	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.580%
43) trans-1,3-Dichloroprop...	8.952	79	22955	37.107	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.220%
47) 2-Hexanone-d5	9.385	63	49158	83.174	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.170%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	80793	42.449	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.900%
66) 1,2-Dichlorobenzene-d4	12.317	152	53443	46.409	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.820%
Target Compounds						
13) Acetone	2.386	43	1779	2.844	ug/L	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
Data File : VX042648.D
Acq On : 27 Jul 2024 19:36
Operator : JC/MD
Sample : P3373-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E08K0

Quant Time: Jul 29 02:46:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jul 29 02:34:37 2024
Response via : Initial Calibration

