

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
 Data File : VX042651.D
 Acq On : 27 Jul 2024 20:45
 Operator : JC/MD
 Sample : VX0728WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK704

Quant Time: Jul 29 03:11:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jul 29 03:03:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	165236	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	146242	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	56292	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	38404	38.256	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.520%
7) Chloroethane-d5	1.648	69	18471	44.652	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.300%
11) 1,1-Dichloroethene-d2	2.294	65	21275	43.013	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.020%
21) 2-Butanone-d5	4.471	46	71322	95.860	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.860%
24) Chloroform-d	5.056	84	100496	45.736	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.480%
26) 1,2-Dichloroethane-d4	5.958	65	67274	47.457	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.920%
32) Benzene-d6	5.971	84	197921	48.764	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.520%
36) 1,2-Dichloropropane-d6	7.306	67	60117	50.047	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.100%
41) Toluene-d8	8.647	98	169965	46.706	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.420%
43) trans-1,3-Dichloroprop...	8.952	79	23141	41.166	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.340%
47) 2-Hexanone-d5	9.385	63	52494	97.743	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.740%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	84765	49.010	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.020%
66) 1,2-Dichlorobenzene-d4	12.323	152	55193	51.880	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072824\
Data File : VX042651.D
Acq On : 27 Jul 2024 20:45
Operator : JC/MD
Sample : VX0728WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK704

Quant Time: Jul 29 03:11:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML072624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jul 29 03:03:42 2024
Response via : Initial Calibration

