

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
 Data File : VX047153.D
 Acq On : 28 Jul 2025 09:58
 Operator : JC/MD
 Sample : VX0728MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0728MBL01

Quant Time: Jul 29 02:11:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	380893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	711791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	673199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	348234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	278066	58.605	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 117.220%#	
35) Dibromofluoromethane	5.391	113	237956	54.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.280%	
50) Toluene-d8	8.647	98	847990	59.581	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 119.160%#	
62) 4-Bromofluorobenzene	11.079	95	358768	58.494	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.980%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
Data File : VX047153.D
Acq On : 28 Jul 2025 09:58
Operator : JC/MD
Sample : VX0728MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0728MBL01

Quant Time: Jul 29 02:11:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

