

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
 Data File : VX047160.D
 Acq On : 28 Jul 2025 13:16
 Operator : JC/MD
 Sample : Q2693-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 29 02:15:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	312397	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	530955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	478257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	245775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	179721	46.183	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.360%
35) Dibromofluoromethane	5.391	113	151842	46.317	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.640%
50) Toluene-d8	8.647	98	464483	43.750	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.500%#
62) 4-Bromofluorobenzene	11.079	95	218448	47.746	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
Data File : VX047160.D
Acq On : 28 Jul 2025 13:16
Operator : JC/MD
Sample : Q2693-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 29 02:15:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

