

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
 Data File : VX047161.D
 Acq On : 28 Jul 2025 13:42
 Operator : JC/MD
 Sample : PB169005TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB169005TB

Quant Time: Jul 29 02:48:09 2025
 Quant Title :
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	392521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	672910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	582759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	296033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	281846	57.642	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 115.280%		
35) Dibromofluoromethane	5.391	113	221915	53.412	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.820%		
50) Toluene-d8	8.647	98	795902	59.152	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 118.300%#		
62) 4-Bromofluorobenzene	11.079	95	301421	51.983	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.960%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	62955	32.140	ug/l	93
18) Methyl Acetate	2.715	43	30442	5.848	ug/l	99
20) Methylene Chloride	2.800	84	24941	4.628	ug/l	93
37) Ethyl Acetate	4.733	43	18732m	2.870	ug/l	
43) Isopropyl Acetate	6.348	43	14899	1.464	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
Data File : VX047161.D
Acq On : 28 Jul 2025 13:42
Operator : JC/MD
Sample : PB169005TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB169005TB

Quant Time: Jul 29 02:48:09 2025
Quant Title :
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

