

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
 Data File : VX047165.D
 Acq On : 28 Jul 2025 15:07
 Operator : JC/MD
 Sample : Q2681-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NYPA-POUCH-SPENT-CARBONDL

Quant Time: Jul 29 02:18:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	322992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	546891	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	497406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	253109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	180413	44.840	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.680%
35) Dibromofluoromethane	5.397	113	154136	45.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.300%
50) Toluene-d8	8.653	98	476198	43.547	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.100%#
62) 4-Bromofluorobenzene	11.079	95	225502	47.852	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	295392	644.263	ug/l	100
16) Acetone	2.392	43	145006	89.965	ug/l	97
18) Methyl Acetate	2.715	43	21774	5.083	ug/l #	81
31) Cyclohexane	5.483	56	13179	1.752	ug/l	94
39) Methylcyclohexane	7.385	83	17259	2.402	ug/l	100
40) Benzene	6.050	78	648253	38.917	ug/l	98
52) Toluene	8.720	92	107067	10.453	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072825\
Data File : VX047165.D
Acq On : 28 Jul 2025 15:07
Operator : JC/MD
Sample : Q2681-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NYPA-POUCH-SPENT-CARBONDL

Quant Time: Jul 29 02:18:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

