

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011145.D
 Acq On : 29 Jul 2019 15:02
 Operator : JC/SP
 Sample : K4018-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU19-MSW0414

Quant Time: Jul 30 05:46:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	290990	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	272109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	100955	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
35) Dibromofluoromethane	5.50	113	91218	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	8.71	98	350388	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.13	95	124223	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
Target Compounds						
16) Acetone	2.45	43	2945	3.401	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011145.D
Acq On : 29 Jul 2019 15:02
Operator : JC/SP
Sample : K4018-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU19-MSW0414

Quant Time: Jul 30 05:46:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

