

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011156.D
 Acq On : 29 Jul 2019 19:19
 Operator : JC/SP
 Sample : K4026-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0681

Quant Time: Jul 30 06:08:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291133	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274617	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126824	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	100620	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
35) Dibromofluoromethane	5.50	113	94091	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
50) Toluene-d8	8.71	98	354519	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	11.13	95	122941	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
Target Compounds						
4) Vinyl Chloride	1.41	62	21365	12.689	ug/l	95
16) Acetone	2.45	43	2102	2.437	ug/l	94
21) trans-1,2-Dichloroethene	3.17	96	1381	0.814	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	10949	5.576	ug/l	91
44) Trichloroethene	7.22	130	1787	0.824	ug/l	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011156.D
Acq On : 29 Jul 2019 19:19
Operator : JC/SP
Sample : K4026-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0681

Quant Time: Jul 30 06:08:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

