

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011167.D
 Acq On : 30 Jul 2019 00:46
 Operator : JC/SP
 Sample : K4026-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0685

Quant Time: Jul 30 06:56:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299088	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	278031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	102919	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
35) Dibromofluoromethane	5.50	113	95998	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
50) Toluene-d8	8.71	98	363673	51.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.13	95	124378	47.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.04%	
Target Compounds						
4) Vinyl Chloride	1.41	62	29532	16.867	ug/l	98
12) 1,1-Dichloroethene	2.38	96	4463	2.684	ug/l	98
16) Acetone	2.45	43	18344	20.453	ug/l	93
21) trans-1,2-Dichloroethene	3.17	96	360113	204.132	ug/l	97
25) 2-Butanone	4.68	43	3666	2.848	ug/l #	87
27) cis-1,2-Dichloroethene	4.60	96	967589	473.920	ug/l	89
44) Trichloroethene	7.22	130	2422582	1087.542	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011167.D
Acq On : 30 Jul 2019 00:46
Operator : JC/SP
Sample : K4026-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0685

Quant Time: Jul 30 06:56:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

