

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011168.D
 Acq On : 30 Jul 2019 01:09
 Operator : JC/SP
 Sample : K4026-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-ER-2019-2

Quant Time: Jul 30 06:58:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	182687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	293124	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279933	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130900	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	102446	52.71	ug/l	0.00
Spiked Amount			Recovery	=	105.42%	
35) Dibromofluoromethane	5.50	113	94445	50.76	ug/l	0.00
Spiked Amount			Recovery	=	101.52%	
50) Toluene-d8	8.71	98	357230	51.19	ug/l	0.00
Spiked Amount			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	11.13	95	125612	48.97	ug/l	0.00
Spiked Amount			Recovery	=	97.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	4972	5.664	ug/l	93
20) Methylene Chloride	2.86	84	18004	9.678	ug/l	91
21) trans-1,2-Dichloroethene	3.18	96	1648	0.954	ug/l #	69
25) 2-Butanone	4.67	43	17113	13.581	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	2180	1.091	ug/l	94
29) Tetrahydrofuran	5.13	42	113332	141.991	ug/l	98
44) Trichloroethene	7.21	130	7179	3.288	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011168.D
Acq On : 30 Jul 2019 01:09
Operator : JC/SP
Sample : K4026-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-ER-2019-2

Quant Time: Jul 30 06:58:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

