

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011173.D
 Acq On : 30 Jul 2019 03:06
 Operator : JC/SP
 Sample : K4021-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-7

Quant Time: Jul 30 07:10:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185423	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	300088	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	298356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	139195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	104936	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
35) Dibromofluoromethane	5.50	113	96266	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	8.71	98	378439	52.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.94%	
62) 4-Bromofluorobenzene	11.13	95	133046	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	7012	7.870	ug/l	95
20) Methylene Chloride	2.86	84	6686	3.541	ug/l	97
30) Chloroform	5.21	83	3766	1.186	ug/l	94
43) Isopropyl Acetate	6.45	43	10267	2.474	ug/l	97
44) Trichloroethene	7.21	130	6279	2.809	ug/l	96
78) n-propylbenzene	11.36	91	10558	1.016	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011173.D
Acq On : 30 Jul 2019 03:06
Operator : JC/SP
Sample : K4021-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-7

Quant Time: Jul 30 07:10:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

