

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011177.D
 Acq On : 30 Jul 2019 04:40
 Operator : JC/SP
 Sample : K4021-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-10

Quant Time: Jul 30 08:19:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	186564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	302025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	295867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137690	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	103327	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	5.49	113	96558	50.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.74%	
50) Toluene-d8	8.71	98	377521	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
62) 4-Bromofluorobenzene	11.13	95	130959	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
Target Compounds						
16) Acetone	2.45	43	8613	9.608	ug/l #	79
20) Methylene Chloride	2.86	84	180906	95.229	ug/l	98
43) Isopropyl Acetate	6.45	43	10224	2.448	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011177.D
Acq On : 30 Jul 2019 04:40
Operator : JC/SP
Sample : K4021-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-10

Quant Time: Jul 30 08:19:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

