

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011179.D
 Acq On : 30 Jul 2019 05:27
 Operator : JC/SP
 Sample : PB121753ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121753ZHE#02

Quant Time: Jul 30 08:23:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	184285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	300968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	293130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	104028	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
35) Dibromofluoromethane	5.50	113	95112	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	8.71	98	376096	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	11.13	95	129590	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
Target Compounds						
16) Acetone	2.44	43	495883	560.027	ug/l	98
20) Methylene Chloride	2.86	84	155165	82.689	ug/l	99
43) Isopropyl Acetate	6.45	43	11583	2.783	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011179.D
Acq On : 30 Jul 2019 05:27
Operator : JC/SP
Sample : PB121753ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121753ZHE#02

Quant Time: Jul 30 08:23:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

