

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011181.D
 Acq On : 30 Jul 2019 06:14
 Operator : JC/SP
 Sample : PB121753ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121753ZHE#04

Quant Time: Jul 30 08:27:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	292496	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	278258	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117380	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	99888	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
35) Dibromofluoromethane	5.50	113	92052	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	8.71	98	366717	52.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.34%	
62) 4-Bromofluorobenzene	11.13	95	120697	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
Target Compounds						
16) Acetone	2.44	43	505871	585.021	ug/l	97
20) Methylene Chloride	2.86	84	152071	82.986	ug/l	99
43) Isopropyl Acetate	6.45	43	11218	2.773	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011181.D
Acq On : 30 Jul 2019 06:14
Operator : JC/SP
Sample : PB121753ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121753ZHE#04

Quant Time: Jul 30 08:27:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

