

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072919\
 Data File : VX011135.D
 Acq On : 29 Jul 2019 10:54
 Operator : JC/SP
 Sample : VX0729WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0729WBL01

Quant Time: Jul 30 05:30:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	192852	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	308355	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	283863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	129857	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	107019	52.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.32%	
35) Dibromofluoromethane	5.49	113	100350	51.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.54%	
50) Toluene-d8	8.71	98	370519	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.13	95	125387	46.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.94%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072919\
Data File : VX011135.D
Acq On : 29 Jul 2019 10:54
Operator : JC/SP
Sample : VX0729WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0729WBL01

Quant Time: Jul 30 05:30:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.00 min

Lab File: VX011135.D

Acq: 29 Jul 2019 10:54

Instrument :

MSVOA_X

ClientSampleId :


VX0729WBL01

Tot Ion:168 Resp: 192852

Ion Ratio Lower Upper

168 100

99 51.6 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 52.162 ug/l

RT: 6.05 min Scan# 804

Delta R.T. -0.01 min

Lab File: VX011135.D


Acq: 29 Jul 2019 10:54

Tgt Ion: 65 Resp: 107019

Ion Ratio Lower Upper

65 100

67 53.0 0.0 107.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX011135.D

Acq: 29 Jul 2019 10:54

Instrument :

MSVOA_X

ClientSampleId :

VX0729WBL01


Tot Ion:114 Resp: 308355

Ion Ratio Lower Upper

114 100

63 18.5 0.0 35.6




88 14.7 0.0 29.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.272 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX011135.D

Acq: 29 Jul 2019 10:54


Tgt Ion:113 Resp: 100350

Ion Ratio Lower Upper

113 100

111 100.0 82.3 123.5



192 22.0 17.4 26.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.474 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX011135.D
Acq: 29 Jul 2019 10:54




Instrument :
MSVOA_X
ClientSampled :
VX0729WBL01


Tot Ion: 98 Resp: 370519
Ion Ratio Lower Upper
98 100
100 63.6 50.6 75.8



#62
4-Bromofluorobenzene
Concen: 46.470 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX011135.D
Acq: 29 Jul 2019 10:54

Tgt Ion: 95 Resp: 125387
Ion Ratio Lower Upper
95 100
174 92.7 0.0 180.6
176 90.1 0.0 173.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX011135.D
Acq: 29 Jul 2019 10:54

Instrument :
MSVOA_X
ClientSampleId :
VX0729WBL01

Tot Ion:117 Resp: 283863
Ion Ratio Lower Upper
117 100
82 50.6 42.6 63.8
119 31.7 25.9 38.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1791
Delta R.T. -0.01 min
Lab File: VX011135.D
Acq: 29 Jul 2019 10:54

Tgt Ion:152 Resp: 129857
Ion Ratio Lower Upper
152 100
115 55.3 39.7 119.1
150 156.9 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

