


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072919\
Data File : VX011133.D
Acq On : 29 Jul 2019 09:09
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Title : SW846 8260
Last Update : Wed Jul 24 05:39:23 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	5252	PASS
75	95	30	60	50.8	15009	PASS
95	95	100	100	100.0	29541	PASS
96	95	5	9	6.6	1954	PASS
173	174	0.00	2	1.5	393	PASS
174	95	50	100	87.6	25877	PASS
175	174	5	9	8.2	2113	PASS
176	174	95	101	96.1	24861	PASS
177	176	5	9	7.6	1893	PASS