

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072920\
 Data File : VX017666.D
 Acq On : 29 Jul 2020 11:07
 Operator : JC/SP
 Sample : VX0729WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0729WBL01

Quant Time: Jul 30 04:47:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	512802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	830019	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	873273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	461850	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	338926	51.45	ug/l	0.00
Spiked Amount			Recovery	=	102.90%	
35) Dibromofluoromethane	5.46	113	289059	52.07	ug/l	0.00
Spiked Amount			Recovery	=	104.14%	
50) Toluene-d8	8.70	98	1040172	51.94	ug/l	0.00
Spiked Amount			Recovery	=	103.88%	
62) 4-Bromofluorobenzene	11.12	95	434994	54.28	ug/l	0.00
Spiked Amount			Recovery	=	108.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072920\
Data File : VX017666.D
Acq On : 29 Jul 2020 11:07
Operator : JC/SP
Sample : VX0729WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0729WBL01

Quant Time: Jul 30 04:47:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

