

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030297.D
 Acq On : 29 Jul 2022 13:02
 Operator : JC/MD
 Sample : N3975-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220727002-01-TRIP-BLANK

Quant Time: Aug 01 05:11:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	268199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169062	55.067	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.140%			
35) Dibromofluoromethane	5.391	113	132759	51.781	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.560%			
50) Toluene-d8	8.653	98	503849	50.746	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.500%			
62) 4-Bromofluorobenzene	11.085	95	174702	47.411	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.820%			
Target Compounds						
16) Acetone	2.386	43	366784	233.350	ug/l	95
30) Chloroform	5.105	83	3824	0.780	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
Data File : VX030297.D
Acq On : 29 Jul 2022 13:02
Operator : JC/MD
Sample : N3975-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220727002-01-TRIP-BLANK

Quant Time: Aug 01 05:11:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

