

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030299.D
 Acq On : 29 Jul 2022 13:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 01 05:12:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	260878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147052	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	153156	50.918	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.840%
35) Dibromofluoromethane	5.397	113	117886	47.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	8.653	98	454015	47.010	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.020%
62) 4-Bromofluorobenzene	11.085	95	160178	44.689	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
Data File : VX030299.D
Acq On : 29 Jul 2022 13:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 01 05:12:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

