

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072922\  
 Data File : VX030301.D  
 Acq On : 29 Jul 2022 14:35  
 Operator : JC/MD  
 Sample : N3965-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SVE-VZC-D4BR-62

Quant Time: Aug 01 05:13:25 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 25 16:49:17 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 149513   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 263337   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 264188   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 148959   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 163136   | 54.286   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 108.580% |
| 35) Dibromofluoromethane    | 5.391  | 113   | 128984   | 51.238   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 102.480% |
| 50) Toluene-d8              | 8.653  | 98    | 485993   | 49.851   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 99.700%  |
| 62) 4-Bromofluorobenzene    | 11.086 | 95    | 168572   | 46.592   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 93.180%  |
| Target Compounds            |        |       |          |          |       |          |
|                             |        |       |          |          |       | Qvalue   |
| 16) Acetone                 | 2.386  | 43    | 5076     | 3.299    | ug/l  | # 79     |
| 17) Carbon Disulfide        | 2.508  | 76    | 5482     | 0.870    | ug/l  | 98       |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072922\  
Data File : VX030301.D  
Acq On : 29 Jul 2022 14:35  
Operator : JC/MD  
Sample : N3965-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SVE-VZC-D4BR-62

Quant Time: Aug 01 05:13:25 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072522W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 25 16:49:17 2022  
Response via : Initial Calibration

