

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030302.D
 Acq On : 29 Jul 2022 14:59
 Operator : JC/MD
 Sample : N3861-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Aug 01 05:13:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	263087	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264060	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146243	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	164984	54.970	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.940%
35) Dibromofluoromethane	5.391	113	127985	50.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	8.653	98	478962	49.177	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.085	95	168747	46.685	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
Data File : VX030302.D
Acq On : 29 Jul 2022 14:59
Operator : JC/MD
Sample : N3861-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Aug 01 05:13:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

