

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030305.D
 Acq On : 29 Jul 2022 16:09
 Operator : JC/MD
 Sample : N3861-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16R

Quant Time: Aug 01 05:14:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145887	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	158801	54.157	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	5.391	113	124840	49.875	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.760%	
50) Toluene-d8	8.653	98	470373	48.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.040%	
62) 4-Bromofluorobenzene	11.085	95	166875	46.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.780%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7135	4.753	ug/l	91
31) Cyclohexane	5.470	56	13856	3.255	ug/l #	77
39) Methylcyclohexane	7.373	83	5781	1.243	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
Data File : VX030305.D
Acq On : 29 Jul 2022 16:09
Operator : JC/MD
Sample : N3861-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16R

Quant Time: Aug 01 05:14:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

