

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030308.D
 Acq On : 29 Jul 2022 17:19
 Operator : JC/MD
 Sample : N3861-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19D

Quant Time: Aug 01 05:16:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	158991	58.263	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.520%
35) Dibromofluoromethane	5.391	113	123626	53.416	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.840%
50) Toluene-d8	8.653	98	458632	51.170	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.340%
62) 4-Bromofluorobenzene	11.085	95	159290	47.887	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.780%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	3086	4.636	ug/l	# 90
16) Acetone	2.380	43	4940	3.536	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	6072	0.836	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
Data File : VX030308.D
Acq On : 29 Jul 2022 17:19
Operator : JC/MD
Sample : N3861-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19D

Quant Time: Aug 01 05:16:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

