



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072922\
 Data File : VX030289.D
 Acq On : 29 Jul 2022 08:57
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Title : SW846 8260
 Last Update : Mon Jul 15 16:49:17 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.0	10704	PASS
75	95	30	60	50.2	26787	PASS
95	95	100	100	100.0	53397	PASS
96	95	5	9	7.4	3976	PASS
173	174	0.00	2	1.1	412	PASS
174	95	50	100	68.1	36352	PASS
175	174	5	9	7.3	2657	PASS
176	174	95	101	95.3	34656	PASS
177	176	5	9	6.9	2376	PASS