

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072925\
 Data File : VX047172.D
 Acq On : 29 Jul 2025 10:19
 Operator : JC/MD
 Sample : VX0729MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0729MBL01

Quant Time: Jul 30 01:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	234873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	435500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	430801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	221028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	176153	60.207	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.420%#	
35) Dibromofluoromethane	5.391	113	145504	54.112	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.220%	
50) Toluene-d8	8.647	98	536842	61.649	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 123.300%#	
62) 4-Bromofluorobenzene	11.079	95	227628	60.658	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 121.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072925\
Data File : VX047172.D
Acq On : 29 Jul 2025 10:19
Operator : JC/MD
Sample : VX0729MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0729MBL01

Quant Time: Jul 30 01:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

