

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011190.D
 Acq On : 30 Jul 2019 13:23
 Operator : JC/SP
 Sample : K4043-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB203-GW-243-245

Quant Time: Jul 31 08:04:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	227273	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	326670	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	288601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136524	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112447	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
35) Dibromofluoromethane	5.50	113	98389	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	8.71	98	382647	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	125092	43.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.52%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.38	96	1049	0.518	ug/l	91
16) Acetone	2.45	43	3899	3.570	ug/l	96
17) Carbon Disulfide	2.57	76	1535	0.296	ug/l #	86
24) 1,1-Dichloroethane	3.70	63	5219	1.449	ug/l	97
44) Trichloroethene	7.22	130	1544	0.635	ug/l	83
64) Tetrachloroethene	9.34	164	1508	0.664	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011190.D
Acq On : 30 Jul 2019 13:23
Operator : JC/SP
Sample : K4043-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB203-GW-243-245

Quant Time: Jul 31 08:04:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

