

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011197.D
 Acq On : 30 Jul 2019 16:06
 Operator : JC/SP
 Sample : K4043-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB203-GW-398-400

Quant Time: Jul 31 08:42:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	219278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	317250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	279985	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	108990	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
35) Dibromofluoromethane	5.50	113	94288	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
50) Toluene-d8	8.71	98	366242	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
62) 4-Bromofluorobenzene	11.13	95	121586	43.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.60%	
Target Compounds						
12) 1,1-Dichloroethene	2.38	96	2740	1.403	ug/l	82
16) Acetone	2.44	43	2905	2.757	ug/l	98
24) 1,1-Dichloroethane	3.70	63	15511	4.464	ug/l	99
27) cis-1,2-Dichloroethene	4.61	96	3357	1.400	ug/l	94
42) 1,2-Dichloroethane	6.20	62	4186	1.508	ug/l	89
44) Trichloroethene	7.21	130	1832	0.775	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011197.D
Acq On : 30 Jul 2019 16:06
Operator : JC/SP
Sample : K4043-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB203-GW-398-400

Quant Time: Jul 31 08:42:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

