

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011205.D
 Acq On : 30 Jul 2019 19:14
 Operator : JC/SP
 Sample : K4026-06DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0674DL

Quant Time: Jul 31 09:08:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211675	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	306283	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	127858	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	105450	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
35) Dibromofluoromethane	5.50	113	88916	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	8.71	98	355309	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.13	95	118769	44.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.64%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	2080	1.048	ug/l #	86
21) trans-1,2-Dichloroethene	3.17	96	1805	0.902	ug/l	91
27) cis-1,2-Dichloroethene	4.60	96	87824	37.932	ug/l	88
40) Benzene	6.15	78	2924	0.379	ug/l	97
65) Chlorobenzene	10.14	112	10774	1.933	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011205.D
Acq On : 30 Jul 2019 19:14
Operator : JC/SP
Sample : K4026-06DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0674DL

Quant Time: Jul 31 09:08:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

