

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011206.D
 Acq On : 30 Jul 2019 19:37
 Operator : JC/SP
 Sample : K4026-07DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-0675DL

Quant Time: Jul 31 09:13:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	207762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	296892	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	270264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	127241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	102660	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
35) Dibromofluoromethane	5.49	113	86988	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
50) Toluene-d8	8.71	98	344664	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	11.13	95	117134	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	1753	0.900	ug/l #	84
21) trans-1,2-Dichloroethene	3.17	96	1817	0.925	ug/l #	74
27) cis-1,2-Dichloroethene	4.59	96	79042	34.782	ug/l	90
40) Benzene	6.13	78	2563	0.343	ug/l #	81
65) Chlorobenzene	10.13	112	9645	1.738	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011206.D
Acq On : 30 Jul 2019 19:37
Operator : JC/SP
Sample : K4026-07DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-0675DL

Quant Time: Jul 31 09:13:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

