

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011207.D
 Acq On : 30 Jul 2019 20:00
 Operator : JC/SP
 Sample : K4048-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWL-C8-GW

Quant Time: Jul 31 09:15:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	323279	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	286817	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112582	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
35) Dibromofluoromethane	5.50	113	96653	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
50) Toluene-d8	8.71	98	373993	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.13	95	121093	42.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.62%	
Target Compounds						
16) Acetone	2.45	43	19562	18.254	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011207.D
Acq On : 30 Jul 2019 20:00
Operator : JC/SP
Sample : K4048-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWL-C8-GW

Quant Time: Jul 31 09:15:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

