

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011208.D
 Acq On : 30 Jul 2019 20:24
 Operator : JC/SP
 Sample : K4048-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWL-C7-GW

Quant Time: Jul 31 09:16:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	209670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303818	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	127241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	106095	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
35) Dibromofluoromethane	5.50	113	90849	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
50) Toluene-d8	8.71	98	351848	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
62) 4-Bromofluorobenzene	11.13	95	115444	43.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.84%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	6272	6.226	ug/l	93
17) Carbon Disulfide	2.57	76	2436	0.508	ug/l	98
19) Methyl tert-butyl Ether	3.20	73	1796	0.299	ug/l #	86
39) Methylcyclohexane	7.46	83	881	0.272	ug/l #	70
73) Isopropylbenzene	11.02	105	18359	2.135	ug/l	99
83) tert-Butylbenzene	11.77	119	4991	0.707	ug/l	92
85) sec-Butylbenzene	11.94	105	15076	1.823	ug/l	79
89) n-Butylbenzene	12.38	91	2484	0.380	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011208.D
Acq On : 30 Jul 2019 20:24
Operator : JC/SP
Sample : K4048-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWL-C7-GW

Quant Time: Jul 31 09:16:29 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

