

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011219.D
 Acq On : 31 Jul 2019 01:27
 Operator : JC/SP
 Sample : PB121808TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121808TB

Quant Time: Jul 31 09:40:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	347442	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	316967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	142802	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	118851	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
35) Dibromofluoromethane	5.50	113	103894	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
50) Toluene-d8	8.71	98	407288	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
62) 4-Bromofluorobenzene	11.13	95	136415	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	
Target Compounds						
16) Acetone	2.45	43	5227	4.494	ug/l	89
20) Methylene Chloride	2.86	84	7460	3.027	ug/l	93
43) Isopropyl Acetate	6.45	43	11878	2.472	ug/l	100
68) m/p-Xylenes	10.37	106	2004	0.471	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011219.D
Acq On : 31 Jul 2019 01:27
Operator : JC/SP
Sample : PB121808TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121808TB

Quant Time: Jul 31 09:40:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

