

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011220.D
 Acq On : 31 Jul 2019 01:51
 Operator : JC/SP
 Sample : PB121803TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121803TB

Quant Time: Jul 31 09:42:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	335524	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	299261	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137138	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	116122	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
35) Dibromofluoromethane	5.50	113	98252	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
50) Toluene-d8	8.71	98	395746	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	11.13	95	132374	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
16) Acetone	2.45	43	5014	4.503	ug/l #	84
20) Methylene Chloride	2.85	84	7509	3.182	ug/l	97
43) Isopropyl Acetate	6.45	43	11733	2.528	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011220.D
Acq On : 31 Jul 2019 01:51
Operator : JC/SP
Sample : PB121803TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121803TB

Quant Time: Jul 31 09:42:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

