

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011221.D
 Acq On : 31 Jul 2019 02:14
 Operator : JC/SP
 Sample : PB121803ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121803ZHE#01

Quant Time: Jul 31 09:43:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230938	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	333298	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	307635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	140232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115494	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
35) Dibromofluoromethane	5.50	113	98538	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	8.71	98	400222	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	11.13	95	135567	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
Target Compounds						
16) Acetone	2.45	43	5257	4.738	ug/l	100
20) Methylene Chloride	2.86	84	7180	3.053	ug/l	93
43) Isopropyl Acetate	6.44	43	11772	2.554	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011221.D
Acq On : 31 Jul 2019 02:14
Operator : JC/SP
Sample : PB121803ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121803ZHE#01

Quant Time: Jul 31 09:43:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

