

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011227.D
 Acq On : 31 Jul 2019 04:35
 Operator : JC/SP
 Sample : PB121803ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121803ZHE#07

Quant Time: Jul 31 09:51:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221639	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	325983	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	294735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133410	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110205	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
35) Dibromofluoromethane	5.50	113	95158	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	8.71	98	385312	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	11.13	95	126420	44.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.64%	
Target Compounds						
16) Acetone	2.45	43	4723	4.435	ug/l	92
20) Methylene Chloride	2.85	84	2366	1.048	ug/l	92
43) Isopropyl Acetate	6.45	43	11395	2.527	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011227.D
Acq On : 31 Jul 2019 04:35
Operator : JC/SP
Sample : PB121803ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121803ZHE#07

Quant Time: Jul 31 09:51:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

