

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011230.D
 Acq On : 31 Jul 2019 05:45
 Operator : JC/SP
 Sample : K4045-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW7-20190726

Quant Time: Jul 31 09:55:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	215867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	314463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	277659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131467	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	108813	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
35) Dibromofluoromethane	5.50	113	94054	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
50) Toluene-d8	8.71	98	366143	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	121348	44.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.20%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011230.D
Acq On : 31 Jul 2019 05:45
Operator : JC/SP
Sample : K4045-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW7-20190726

Quant Time: Jul 31 09:55:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

