

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011232.D
 Acq On : 31 Jul 2019 06:32
 Operator : JC/SP
 Sample : K3621-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-01-072919

Quant Time: Jul 31 09:59:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	319168	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	297820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	137647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110137	46.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.24%	
35) Dibromofluoromethane	5.50	113	95099	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
50) Toluene-d8	8.71	98	380256	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.13	95	128616	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
Target Compounds						
16) Acetone	2.45	43	7276	6.819	ug/l	98
20) Methylene Chloride	2.85	84	88029	38.932	ug/l	99
30) Chloroform	5.21	83	3753	0.987	ug/l	97
43) Isopropyl Acetate	6.45	43	10531	2.386	ug/l	98
95) Naphthalene	13.83	128	12071	1.367	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011232.D
Acq On : 31 Jul 2019 06:32
Operator : JC/SP
Sample : K3621-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-01-072919

Quant Time: Jul 31 09:59:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

