

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073019\
 Data File : VX011236.D
 Acq On : 31 Jul 2019 08:05
 Operator : JC/SP
 Sample : K4044-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5

Quant Time: Jul 31 10:10:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220479	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	323204	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	294517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110585	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	5.50	113	95597	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
50) Toluene-d8	8.71	98	386309	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.13	95	126605	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
Target Compounds						
16) Acetone	2.45	43	5330	5.031	ug/l	98
20) Methylene Chloride	2.86	84	9041	4.027	ug/l	96
43) Isopropyl Acetate	6.45	43	10012	2.240	ug/l	98
44) Trichloroethene	7.21	130	13253	5.506	ug/l	91
95) Naphthalene	13.83	128	5449	0.624	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073019\
Data File : VX011236.D
Acq On : 31 Jul 2019 08:05
Operator : JC/SP
Sample : K4044-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5

Quant Time: Jul 31 10:10:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

