

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073020\
 Data File : VX017678.D
 Acq On : 30 Jul 2020 12:54
 Operator : JC/SP
 Sample : L3514-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PK-10S

Quant Time: Jul 31 11:35:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	468747	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	702267	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	825956	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	414353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	297511	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
35) Dibromofluoromethane	5.46	113	263710	56.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.30%	
50) Toluene-d8	8.70	98	867216	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.12	95	413558	60.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.98%	
Target Compounds						
16) Acetone	2.42	43	6597	2.489	ug/l	91
93) 1,2,4-Trichlorobenzene	13.63	180	3115	0.340	ug/l	94
94) Hexachlorobutadiene	13.77	225	1809	0.429	ug/l	90
95) Naphthalene	13.82	128	6214	0.227	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073020\
Data File : VX017678.D
Acq On : 30 Jul 2020 12:54
Operator : JC/SP
Sample : L3514-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PK-10S

Quant Time: Jul 31 11:35:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

