

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073020\
 Data File : VX017679.D
 Acq On : 30 Jul 2020 13:17
 Operator : JC/SP
 Sample : L3514-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PK-10D

Quant Time: Jul 31 11:39:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	499323	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	837525	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	877470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	472689	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	342087	53.33	ug/l	0.00
Spiked Amount						
			Recovery	=	106.66%	
35) Dibromofluoromethane	5.47	113	286958	51.23	ug/l	0.00
Spiked Amount						
			Recovery	=	102.46%	
50) Toluene-d8	8.70	98	1037218	51.33	ug/l	0.00
Spiked Amount						
			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	11.12	95	447887	55.38	ug/l	0.00
Spiked Amount						
			Recovery	=	110.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	6442	2.282	ug/l	93
30) Chloroform	5.18	83	88797	8.761	ug/l	100
44) Trichloroethene	7.20	130	3806	0.566	ug/l	98
47) Bromodichloromethane	7.87	83	1983	0.218	ug/l #	71
64) Tetrachloroethene	9.32	164	3505	0.445	ug/l #	70
65) Chlorobenzene	10.12	112	19801	1.043	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073020\
Data File : VX017679.D
Acq On : 30 Jul 2020 13:17
Operator : JC/SP
Sample : L3514-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PK-10D

Quant Time: Jul 31 11:39:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

