

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073020\
 Data File : VX017684.D
 Acq On : 30 Jul 2020 15:34
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 31 06:32:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	466112	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	788831	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	831448	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	460384	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	324528	54.20	ug/l	0.00
Spiked Amount						
			Recovery	=	108.40%	
35) Dibromofluoromethane	5.47	113	268234	50.84	ug/l	0.00
Spiked Amount						
			Recovery	=	101.68%	
50) Toluene-d8	8.70	98	977270	51.35	ug/l	0.00
Spiked Amount						
			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	11.12	95	429410	56.38	ug/l	0.00
Spiked Amount						
			Recovery	=	112.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073020\
Data File : VX017684.D
Acq On : 30 Jul 2020 15:34
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
IBLK

Quant Time: Jul 31 06:32:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

