

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
 Data File : VX023508.D
 Acq On : 29 Jul 2021 13:08
 Operator : JC/MD
 Sample : VX0730WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0730WBL01

Quant Time: Jul 30 03:40:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	168697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	281774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90958	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109944	52.246	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.500%
35) Dibromofluoromethane	5.391	113	88652	50.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.640%
50) Toluene-d8	8.653	98	329590	50.919	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.840%
62) 4-Bromofluorobenzene	11.085	95	101556	46.140	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
Data File : VX023508.D
Acq On : 29 Jul 2021 13:08
Operator : JC/MD
Sample : VX0730WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0730WBL01

Quant Time: Jul 30 03:40:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

