

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
 Data File : VX023511.D
 Acq On : 29 Jul 2021 14:23
 Operator : JC/MD
 Sample : PB138025TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138025TB

Quant Time: Jul 30 03:42:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	169564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	286996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110585	52.282	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.560%	
35) Dibromofluoromethane	5.397	113	90771	50.585	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.160%	
50) Toluene-d8	8.653	98	346096	52.496	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.000%	
62) 4-Bromofluorobenzene	11.085	95	112096	50.002	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.000%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	7792	7.510	ug/l	93
20) Methylene Chloride	2.794	84	9198	4.557	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
Data File : VX023511.D
Acq On : 29 Jul 2021 14:23
Operator : JC/MD
Sample : PB138025TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138025TB

Quant Time: Jul 30 03:42:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

