

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
 Data File : VX023512.D
 Acq On : 29 Jul 2021 14:47
 Operator : JC/MD
 Sample : PB138025ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138025ZHE#02

Quant Time: Jul 30 03:42:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	162977	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	273631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103754	51.035	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.080%
35) Dibromofluoromethane	5.397	113	84588	49.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	8.653	98	328626	52.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.560%
62) 4-Bromofluorobenzene	11.085	95	103409	48.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
16) Acetone	2.386	43	7436	7.456	ug/l #	84
20) Methylene Chloride	2.794	84	9346	4.818	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
Data File : VX023512.D
Acq On : 29 Jul 2021 14:47
Operator : JC/MD
Sample : PB138025ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138025ZHE#02

Quant Time: Jul 30 03:42:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

