

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
 Data File : VX023513.D
 Acq On : 29 Jul 2021 15:11
 Operator : JC/MD
 Sample : PB138025ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138025ZHE#03

Quant Time: Jul 30 03:42:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	161504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	270801	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104074	51.659	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.320%	
35) Dibromofluoromethane	5.397	113	86593	51.143	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.280%	
50) Toluene-d8	8.653	98	326939	52.556	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.120%	
62) 4-Bromofluorobenzene	11.085	95	100088	47.316	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.640%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	7128	7.213	ug/l	# 87
20) Methylene Chloride	2.794	84	9803	5.100	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
Data File : VX023513.D
Acq On : 29 Jul 2021 15:11
Operator : JC/MD
Sample : PB138025ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138025ZHE#03

Quant Time: Jul 30 03:42:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

